

**A HYBRID HUMAN SENSORY MIMICKING
APPROACH: AN INTEGRATED E-NOSE AND
E-TONGUE SYSTEM**

NAZIFAH BINTI AHMAD FIKRI

© This item is protected by original copyright

**UNIVERSITI MALAYSIA PERLIS
2012**



**A Hybrid Human Sensory Mimicking Approach:
An Integrated E-Nose and E-Tongue System**

by

**Nazifah Binti Ahmad Fikri
(0630610141)**

A thesis submitted
In fulfilment of the requirement for the degree of
Master of Science Mechatronic Engineering

**School of Mechatronic Engineering
UNIVERSITI MALAYSIA PERLIS**

2012

UNIVERSITI MALAYSIA PERLIS

DECLARATION OF THESIS

Author's full name : NAZIFAH BINTI AHMAD FIKRI
Date of birth : 30 MARCH 1983
Title : Hybrid Human Sensory Mimicking Approach: An Integrated E-Nose and.....
E-Tongue.....
.....
Academic Session :

I hereby declare that the thesis becomes the property of Universiti Malaysia Perlis (UniMAP) and to be placed at the library of UniMAP. This thesis is classified as :


- CONFIDENTIAL** (Contains confidential information under the Official Secret Act 1972)*
 RESTRICTED (Contains restricted information as specified by the organization where research was done)*
 OPEN ACCESS I agree that my thesis is to be made immediately available as hard copy or on-line open access (full text)

I, the author, give permission to the UniMAP to reproduce this thesis in whole or in part for the purpose of research or academic exchange only (except during a period of _____ years, if so requested above).

Certified by:



SIGNATURE



SIGNATURE OF SUPERVISOR

830330-03-5330
(NEW IC NO. / PASSPORT NO.)

ASSOC. PROF. DR. ABDUL HAMID ADOM
NAME OF SUPERVISOR

Date : _____

Date : _____

Assoc. Prof. Dr. Abdul Hamid Adom
Dean
School Of Mechatronic Engineering
Universiti Malaysia Perlis
Email: ahamad@unimap.edu.my

NOTES : * If the thesis is CONFIDENTIAL or RESTRICTED, please attach with the letter from the organization with period and reasons for confidentiality or restriction.

ACKNOWLEDGEMENT

I would like to express my gratitude to all those who gave me the possibility to complete this thesis.

To my supervisors, Assoc. Prof. Dr. Abdul Hamid Adom and Mr. Mohd Hafiz Fazalul Rahiman.

To my examiners: Assoc. Prof. Dr. Hazry Desa (UniMAP) and Prof. Dr. Jasmy Yunus (UTM).

To Prof. Dr Ali Yeon Md. Shakaff, who always comes with great ideas and good advices.

To Sensors and Applications Group members, Prof. Dr. Mohd Noor Ahmad, Mr. Abu Hassan Abdullah, Mr. Ammar Zakaria, Mr. Erdi Sulino and others.

Special thanks to ma (Norisah Mohammad) and abah (Ahmad Fikri Ismail) who always motivate me, and also to my husband (Anuar Ibrahim Zainal Abidin) and family for their love and support.

Last but not least to Akid Iman, this is for you, son.

TABLE OF CONTENTS

TITLE	PAGE
ACKNOWLEDGEMENTS	ii
TABLE OF CONTENTS	iii
LIST OF FIGURES	vii
LIST OF TABLES	ix
LIST OF ABBREVIATIONS	x
ABSTRAK	xi
ABSTRACT	xii
CHAPTER 1 INTRODUCTION	
1.1 Introduction.....	1
1.2 Research Objectives.....	3
1.3 Research Scopes.....	3
1.4 Thesis Outlines.....	4
CHAPTER 2 LITERATURE REVIEW	
2.1 Human and Senses.....	6
2.2 Human Sensory Sub-Systems.....	8
2.2.1 Olfaction System.....	10
2.2.2 Gustation System.....	13
2.2.3 Combination of the Human Senses.....	15
2.3 Artificial Sensing System.....	15
2.3.1 Sensor Array for Global Sensing.....	18
2.3.2 Electronic Nose (E-Nose).....	19
2.3.2.1 Gas Sensor.....	21
2.3.2.2 Metal Oxide Gas Sensor (MOS).....	23
2.3.2.3 Recent Work on E-Nose.....	27
2.3.3 Electronic Tongue (E-Tongue).....	29
2.3.3.1 Taste Sensor.....	30
2.3.3.2 Ion-Selective Electrodes.....	31
2.3.3.3 Recent Work on E-Tongue.....	34
2.3.4 Hybrid Systems.....	35
2.4 Data Processing.....	36
2.5 Pre-Processing.....	37
2.5.1 Data Selection.....	38
2.5.1.1 Feature Selection.....	38
2.5.1.2 Feature Extraction.....	39
2.5.2 Dimension Reduction.....	40
2.6 Normalisation.....	41
2.7 Pattern Recognition.....	42
2.7.1 Principal Component Analysis.....	43
2.7.2 Artificial Neural Network.....	44
2.8 Data Fusion.....	47

2.9	System Integration Issues.....	52
2.10	Target Applications.....	55
2.10.1	Agriculture Application.....	56
2.10.2	Environmental Application.....	57
2.10.3	Beverages Representing the Food Industry.....	58
2.11	Discussions.....	59

CHAPTER 3 INITIAL INVESTIGATIONS AND SUB-SYSTEM DEVELOPMENT AND TESTING

3.1	Electronic Nose Sub-System.....	61
3.1.1	Gas Sensor Selection.....	61
3.1.2	Figaro MOS Gas Sensors Basic Operating Procedures.....	63
3.1.3	Basic Electronic Nose Sub-System Operating Procedures.....	65
3.1.4	Electronic Nose Sub-System Classification Evaluation (Food Essences Classification).....	67
3.1.4.1	Electronic Nose Sub-System Sample Odour Profiling and Principal Component Analysis Results.....	70
3.1.4.2	Electronic Nose Sub-System Classification Results.....	72
3.1.5	Comparison of Performance between Commercial E-Nose (Cyranose 320) and In-House E-Nose.....	74
3.1.5.1	Sample Preparation and Data Collection.....	75
3.1.5.2	Data Profile for <i>Ganoderma boninense</i>	75
3.1.5.3	Classification using ANN.....	77
3.1.6	Electronic Nose Sub-System Development and Testing Summary.....	78
3.2	Electronic Tongue Sub-System.....	79
3.2.1	Taste Sensor Selection.....	79
3.2.2	Ion-Selective Sensors Basic Operating Procedures.....	81
3.2.3	Basic Electronic Tongue Sub-System Operating Procedures..	82
3.2.4	Electronic Tongue Sub-System Classification Evaluation (Basic Taste Classification).....	85
3.2.4.1	Electronic Tongue Sub-System Profiling and Principal Component Analysis (PCA) Results.....	86
3.2.4.2	Electronic Tongue Sub-System Classification Results.....	87
3.2.5	Electronic Tongue Sub-System Development and Testing Summary.....	89
3.3	Discussions.....	90

CHAPTER 4 A-HUMANSSENS PROTOTYPE DEVELOPMENT

4.1	A-HumanSens Prototyping.....	91
4.2	The Concept of Integration Multiple Modalities in a Single System.....	92
4.3	Hardware Development.....	94
4.3.1	Signal Conditioning Circuits.....	94
4.3.1.1	Electronic Nose Sub-System Signal Conditioning Circuit.....	95

4.3.1.2	Electronic Tongue Sub-System Signal Conditioning Circuit.....	96
4.3.2	A-HumanSens Main Controller Board.....	98
4.3.3	Prototype Part Fabrication.....	100
4.3.3.1	Sensor Array Holder Design.....	101
4.3.3.2	Rapid Prototyping.....	102
4.4	Software Development.....	106
4.4.1	Graphic User Interface.....	107
4.4.2	Data Processing, Data Fusion and Pattern Recognition.....	109
4.5	System Integration and Testing.....	110
4.5.1	Standard Operating Procedures.....	111
4.5.1.1	Sample Preparation.....	111
4.5.1.2	System Calibration and Preheating.....	112
4.6	Discussions and Summary.....	117

CHAPTER 5 A-HUMANSENS PROTOTYPE TESTING AND EVALUATION

5.1	Sample Applications of Hybrid Sensing System.....	119
5.2	Methodology.....	119
5.3	Sample Preparation.....	120
5.3.1	Agricultural Application Example.....	120
5.3.2	Environmental Application Example.....	122
5.3.3	Beverages Classification Application Example.....	124
5.4	System Setup.....	125
5.5	Data Collection.....	126
5.6	Data Processing.....	126
5.6.1	Fusion Technique.....	126
5.6.2	Data Selection.....	129
5.6.3	Dimension Reduction (Sensors Selection).....	130
5.7	Classification.....	133
5.7.1	Principal Component Analysis (PCA).....	134
5.7.2	Artificial Neural Network (ANN).....	136
5.8	Discussions.....	138

CHAPTER 6 RESULTS AND DISCUSSIONS

6.1	Agricultural Application Experiment.....	140
6.1.1	Sensors Selection for Java Teas Classification.....	140
6.1.2	PCA Results for Java Teas Classification.....	142
6.1.3	ANN Results for Java Teas Classification.....	146
6.2	Environmental Application Experiment.....	148
6.2.1	Sensors Selection for Water Quality Monitoring.....	148
6.2.2	PCA Results for Water Quality Monitoring.....	150
6.2.3	ANN Results for Water Quality Monitoring.....	153
6.3	Beverages Application Experiment.....	156
6.3.1	Sensors Selection for Bottled Mineral and Drinking Water Classification.....	156
6.3.2	PCA Results for Bottled Mineral and Drinking Water Classification.....	158

6.3.3	ANN Results for Bottled Mineral and Drinking Water Classification.....	162
6.4	Discussions.....	162
CHAPTER 7 CONCLUSIONS		
7.1	Conclusions.....	164
7.2	Problems and Obstacles.....	166
7.3	Suggestions for Further Work.....	167
REFERENCES		169
APPENDICES		
	Appendix A	178
	Appendix B	179

© This item is protected by original copyright

LIST OF FIGURES

FIGURES	TITLE	PAGE
Figure 2.1	Organisation of the nervous system in human body	7
Figure 2.2	The connection between afferent neurons and efferent neurons	8
Figure 2.3	The system analogy between human sensing system and artificial sensing system	18
Figure 2.4	Cyranose 320	23
Figure 2.5	TGS Figaro gas sensor structure	24
Figure 2.6	The potential barrier in the absence of gases	25
Figure 2.7	The potential barrier in the presence of gases	26
Figure 2.8	Plot of sensor response characteristic with time	39
Figure 2.9	Basic neuron model showing the components and connections	45
Figure 2.10	MLP architecture	46
Figure 2.11	Flow of data processing for the low level fusion	49
Figure 2.12	Flow of data processing for the high level fusion (or intermediate level fusion)	50
Figure 2.13	Flow of data processing for the high level fusion	51
Figure 3.1	Comparison of the plots of output signals for MOS gas sensors when (a) not heated, (b) heated for one hour	64
Figure 3.2	Sensors response for heating (a) without purging, (b) while purging	64
Figure 3.3	Basic setup for dynamic sampling method	66
Figure 3.4	PCA results for two types of water using e-nose (a) without purging, (b) with purging procedure conducted	67
Figure 3.5	Vial contains 10ml of rose essence	68
Figure 3.6	GUI for e-nose showing the sensor output plots	69
Figure 3.7	Radial data profile for food essences	70
Figure 3.8	PCA results for food essences	71
Figure 3.9	ANN target and output result for food essences	73
Figure 3.10	Sample of <i>Ganoderma boninense</i> in a sealed glass beaker	75
Figure 3.11	Data profile of <i>Ganoderma boninense</i> and ambient odour obtained using Cyranose 320	76
Figure 3.12	Data profile of ambient and <i>Ganoderma boninense</i> odour obtained using in-house e-nose	77
Figure 3.13	Ion-selective sensors response (a) without reference electrode, (b) with reference electrode	81
Figure 3.14	Clustering of the samples using PCA of ion-selective sensors in different dipping level	83
Figure 3.15	Sample clustering using PCA for measurement conducted (a) with bubbler, (b) without using bubbler	84
Figure 3.16	Data profile of five basic tastes provided by the ion-selective sensors responses	86
Figure 3.17	PCA result for basic taste samples using ion-selective sensors	87

Figure 3.18	ANN target and output result for basic taste using e-tongue	88
Figure 4.1	Process flow for A-HumanSens	93
Figure 4.2	Signal conditioning circuit for TGS 2620	95
Figure 4.3	Voltage follower circuit for the ISE sensors	97
Figure 4.4	A-HumanSens basic descriptions	99
Figure 4.5	Pololu USB-to-serial adapter	99
Figure 4.6	The schematic design of the sensor arrays of the A-HumanSens showing the odour capturing module (OCM) for the MOS gas sensors and ion-selective sensors array	100
Figure 4.7	Exploded view of A-HumanSens	103
Figure 4.8	Sensors array chambers for A-HumanSens	104
Figure 4.9	Cross section area for A-HumanSens	105
Figure 4.10	Complete design of A-HumanSens	106
Figure 4.11	A-HumanSens graphic user interface (GUI)	107
Figure 4.12	Flow chart of function of GUI	108
Figure 4.13	Schematics of basic experimental setup	115
Figure 4.14	Basic experimental setup	116
Figure 5.1	Flow chart of low level fusion method for A-HumanSens	128
Figure 5.2	Flow chart of intermediate level fusion method for A-HumanSens	127
Figure 5.3	Flow chart of dimension reduction	133
Figure 5.4	Flow chart of the Principal Component Analysis	135
Figure 5.5	ANN activation function	137
Figure 6.1	PCA plots for Java teas analysis using e-nose	143
Figure 6.2	PCA plots for Java teas analysis using e-tongue	144
Figure 6.3	PCA plots for Java teas analysis using hybrid system with low level fusion	145
Figure 6.4	PCA plots for Java teas analysis using hybrid system with intermediate level fusion	146
Figure 6.5	PCA plots for water quality monitoring using e-nose	150
Figure 6.6	PCA plots for water quality monitoring using e-tongue	151
Figure 6.7	PCA plots for water quality monitoring using hybrid system with low level fusion	152
Figure 6.8	PCA plots for water quality monitoring using hybrid system with intermediate level fusion	153
Figure 6.9	PCA plots for bottled mineral and drinking water classification using e-nose	158
Figure 6.10	PCA plots for bottled mineral and drinking water classification using e-tongue	159
Figure 6.11	PCA plots for bottled mineral and drinking water classification using hybrid system with low level fusion	160
Figure 6.12	PCA plots for bottled mineral and drinking water classification using hybrid system with high level fusion	161

LIST OF TABLES

TABLES	TITLE	PAGE
Table 2.1	Colour, odour and toxicity of elements and common compounds that are gases at ordinary temperatures and pressures	12
Table 2.2	Commercial companies that manufactured sensor-based electronic noses and related instruments in 2003	22
Table 2.3	Baseline correction formula where R_s = steady-state data and R_o = baseline data	40
Table 3.1	List of Figaro MOS gas sensors used and their corresponding sensitivity	62
Table 3.2	Number of training, validation and classification data for each sample	72
Table 3.3	Percentage of classification accuracy results for food essences	74
Table 3.4	Target used in the ANN training	78
Table 3.5	The ANN testing results with percentage of accuracy	78
Table 3.6	List of ion-selective sensors used in the e-tongue sub-system with their sensitivity and concentration range	80
Table 3.7	Samples for basic taste experiment	85
Table 3.8	Percentage of classification accuracy results for basic taste	89
Table 4.1	Load resistance (R_L) used for Figaro gas sensors	96
Table 4.2	GUI button description	109
Table 4.3	Standard suggestion calibration solution for ISE sensors	113
Table 5.1	Java teas classification samples	121
Table 5.2	List of water sources and their pH values	123
Table 5.3	List of bottled mineral and drinking water samples	124
Table 5.4	Mineral concentration in samples (in mg/L)	125
Table 6.1	Tabulated mean correlation coefficients for Java teas analysis using low level and intermediate level fusion. The higher the mean correlation coefficient means that the sensors are linearly correlated	141
Table 6.2	The ANN result for Java teas classification	147
Table 6.3	Tabulated mean correlation coefficients for water quality monitoring using low level and intermediate level fusion. The higher the mean correlation coefficient means that the sensors are linearly correlated	149
Table 6.4	The ANN result for water quality monitoring classification (according to type of sources)	154
Table 6.5	The ANN result for water quality monitoring classification (according to pH value)	155
Table 6.6	Tabulated mean correlation coefficients for bottled mineral and drinking water classification using low level and intermediate level fusion. The higher the mean correlation coefficient means that the sensors are linearly correlated	157
Table 6.7	The ANN results for bottled mineral and drinking water classification	162

LIST OF ABBREVIATIONS

ANN	Artificial Neural Network
BPF	Band Pass Filter
BPNN	Back-Propagation Neural Network
CAD	Computer Aided Design
CCTV	Closed Circuit Television
CNS	Central Nervous System
CP	Conducting Polymer
CPU	Central Processing Unit
FET	Field Effect Transistor
GC	Gas Chromatography
GC-MS	Gas Chromatography-Mass Spectrometry
HCA	Hierarchical Clustering Analysis
HPF	High Pass Filter
ISE	Ion-Selective Electrode
LPF	Low Pass Filter
MLP	Multilayer Perceptron
MOS	Metal Oxide Semiconductor
MS	Mass Spectrometry
MSG	Monosodium Glutamic Acid
NASA	National Aeronautics and Space Administration
OCM	Odour Capturing Module
PCA	Principal Component Analysis
pH	Potential of Hydrogen
PNS	Peripheral Nervous System
RBFNN	Radial Basis Function Neural Network
RP	Rapid Prototyping
QMB	Quartz Crystal Microbalance
SAW	Surface Acoustic Wave
SOMNN	Self-Organising Map Neural Network
SOP	Standard Operating Procedure
TGS	Taguchi Gas Sensor

ABSTRAK

Pendekatan Sistem Hibrid Penderia Manusia Tiruan : Integrasi Sistem Penderia Bau dan Penderia Rasa

Bau dan rasa adalah dua daripada lima penderia manusia yang sama dengan mamalia yang lain. Kedua-dua penderia ini biasanya digunakan bersama untuk melengkapkan persepsi otak manusia terhadap rasa. Keadaan ini menjadi pemangkin kepada kajian dalam gabungan data daripada beberapa sistem penderia buatan seperti penderia bau dan penderia rasa. Walau bagaimanapun, gabungan data yang dijalankan oleh kajian sebelum ini adalah berasaskan sistem modaliti-tunggal yang berasingan. Berikut dibentangkan pembangunan sistem hibrid yang menggabungkan penderia rasa buatan dan penderia bau buatan di dalam sistem tunggal. Kedua-dua sub-sistem menggunakan komponen-komponen yang sedia ada dan dibangunkan menggunakan teknik *rapid prototyping*. Sistem hibrid ini menggabungkan dua kumpulan penderia yang terdiri daripada penderia gas MOS dan elektrod *ion-selective*. Ia juga terdiri daripada unit pengumpulan isyarat data dan perisian pengcaman yang digunakan menerusi komputer. Sistem ini adalah berdasarkan analisis kualiti yang menyamai sistem penderia manusia dengan menggunakan *Principal Component Analysis (PCA)* dan *Artificial Neural Network (ANN)*. Tiga jenis ujian dijalankan mewakili aplikasi-aplikasi di dalam bidang pertanian, alam sekitar dan pengeluaran makanan. Prestasi sistem modaliti-tunggal telah dibandingkan dengan sistem hibrid ini. Keputusan ujian yang dijalankan menunjukkan sistem hibrid memberikan prestasi yang lebih baik berbanding kedua-dua subsistem tunggal, apabila menggunakan kaedah gabungan data yang bersesuaian, dan mampu mencapai ketepatan sehingga 98.67%. Ini telah membuktikan bahawa sistem modaliti-pelbagai yang lebih mirip kepada sistem penderia manusia memberikan prestasi yang lebih baik berbanding sistem modaliti-tunggal dalam membezakan sampel.

ABSTRACT

A Hybrid Human Sensory Mimicking Approach: An Integrated E-Nose and E-Tongue System

Taste and smell are two of the five human senses that are common among mammals. These two senses are usually used together to make up the brain's perception of flavour. This has led to the study of data fusion of multiple artificial sensory systems such as electronic nose and electronic tongue. However, the data fusions performed by these studies are based on separate single-modality systems. Presented is the development of a hybrid system which combines an electronic nose and electronic tongue in a single system. Both sub-systems use off-the-shelf components and were developed using rapid prototyping techniques. The hybrid system combines two sensor arrays of MOS gas sensors and ion-selective electrodes. It also consists of a signal-collecting unit and pattern recognition software applied to a computer. The system uses qualitative analysis which is similar to the human sensory system, implementing Principal Component Analysis (PCA) and Artificial Neural Network (ANN). Three tests were performed representing agricultural, environmental and food production applications. The performance of the single-modality systems was compared to the hybrid system. The results show that the hybrid system performed better than both single sub-systems when appropriate fusion methods were used, and able to achieve up to 98.67% accuracy. This proved that the multi-modality system performed better in sample discrimination than a single-modality system which mimics more closely the human sensory system.

CHAPTER 1

INTRODUCTION

1.1 Introduction

Olfaction (smell) and taste (gustation) are two out of five basic human senses. These two basic senses aid human in daily life. They help to distinguish ripe fruits, bad foods and even used for safety purpose such as the detection of gaseous leakage or burning. Study proved that both senses are related to each other where the human gustatory system cannot perform well if the olfactory system is blocked.

The smell and taste sensation also being used in industrial activity such as in tea and wine classification. However, these human expertise are normally affected by environment and their health and emotion. Therefore, there are many analytical instruments that were designed to replace human expertise.

The analytical instrument such as Gas Chromatography (GC), Mass Spectrometry (MS), Optical Spectrometry and GC-MS are normally used in quantitative analysis. However, there are some disadvantages of using these analytical instruments. The bulky size limits the usage of instrument to laboratory analysis only. These instruments are also expensive and they need an expert in order for these instruments to perform well.

Recently, there are artificial devices that mimic human olfaction and gustation system. They are known as electronic nose (e-nose) and an electronic tongue (e-tongue). These devices more resembling the human olfaction and gustation systems compared to the analytical instruments that have been mention above which normally used in qualitative analysis.

These systems normally consist of an array of sensors with signal handling system, data processing and pattern recognition analysis. The e-nose and e-tongue are being differentiated by their sensors arrays which are gas sensors for e-nose and taste sensors for e-tongue. Both types of sensors work in different type samples. The gas sensors work with the presence of specific gas while the taste sensor work with the presence of specific ion (refer Chapter 2 for further discussion).

It also has been proved that the e-nose and e-tongue can be applied in various applications, from environmental and agricultural industry to medical and other commercial industry (D'Amico, 2000). As human olfaction and gustation sensory system will perform better when both in good condition, thus the combination e-nose and e-tongue which mimic these systems are expected to perform better than the single system.

1.2 Research Objectives

The objectives of this research are:

- i) to develop an in-house hybrid artificial sensory system which combines an e-nose and e-tongue.
- ii) to define the standard operating procedures (SOP) which will take into account of working principles for both e-nose and e-tongue and the specific sensors.
- iii) to investigate the different data fusion techniques (data fusion levels).
- iv) to study and prove whether multi-modality system perform better than single-modality system in samples discrimination and recognition. The hybrid system that combines e-nose and e-tongue is expected to perform better than the single system, which is as similar to the bonding of human olfaction and gustation system.

1.3 Research Scopes

The research work consists of five main scopes which are:

- i) Development of in-house hybrid system which combines e-nose and e-tongue.
- ii) Develop the standard operating procedure (SOP) of the hybrid system.
- iii) Comparison of performance of single-modality system (e-nose and e-tongue) with the multi-modality system (hybrid system).

1.4 Thesis Outline

This thesis is organised into seven chapters. The introduction of this research is presented in Chapter 1. It describes the basic preface of the e-nose and e-tongue. This chapter also explains the objectives of this research as well as the thesis outlines.

Chapter 2 presents the literature review of both sub-systems that will be used in this research. It discusses the general fact of human senses, specifically the olfaction (smell) and gustation (taste) and also the e-nose and e-tongue sub-systems. It also describes the data processing, pattern recognition and data fusion technique that will be used to analyse the data obtained from both sub-systems.

Chapter 3 will discuss the initial study of both sub-systems which includes the sensors selection and the basic operating procedure of all sensors used. The evaluation of a single-modal system also will be presented in this chapter. Also, comparison of performance between in-house and commercial e-nose will be discussed in this chapter.

The development of the combined system will be discussed in Chapter 4. It will include the hardware and software development and the system integration and testing which consists of the standard operating procedure of the combined system.

Chapter 5 will discuss the testing and evaluation of the hybrid system. It includes the system and sample preparation, list of samples used and methods of data analysis.

The results and discussion in Chapter 6 will evaluate the performance of the system. Results from e-nose and e-tongue are compared with the hybrid system which comes from two different fusion methods.

Finally, Chapter 7 will conclude the possibility applications of the combined system. It will also describe the problems faced in this research and suggestions for future work.

© This item is protected by original copyright

CHAPTER 2

LITERATURE REVIEW

This chapter discusses the human sensory systems, and also the related research in the area of electronic sensing systems inspired by the former, and the implementation of data fusion. These literatures illustrate the possibilities in the area of the application of bio-mimicking sensing, and how the manipulation of the data from several single sensing or modality systems can improve their performance. These form the basis of the project, and used as the guidance on the development of the system and also to perform analysis.

The human sensory systems are described, followed by the electronic or artificial counterparts, covering the types of systems available, the choice of sensors as well as their applications. The data processing techniques are also discussed. Single systems are described, followed by the implementation of data fusion, and finally the idea of hybrid system.

2.1 Human and Senses

The human nervous system is a complex communication network system. It is composed of a network of cells throughout the body which receives the information about the internal and external environments. After the information is assessed, a signal is then send to an organ that will react with an appropriate

response. It is one of major systems that used to regulate the entire processes in the body. Hence, this nervous system can be the most important system in human body.

The nervous system consists of two major parts which are the central nervous system (CNS) and the peripheral nervous system (PNS). The CNS includes the brain and the spinal cord while the PNS made up of all of the nervous tissue outside the CNS, as shown in Figure 2.1.




Figure 2.1 Organisation of the nervous system in human body (Standley, 2011).

The brain is an organ, consisting of nervous tissues in the skull. Impulses (signals) from sensory organs and general body control are evaluated in the brain. The spinal cord extends from the brain to the peripheral nerves through the backbone. The nervous tissue consists of bundles of cells that work as circuits for the signals transmissions (Gaudin, 1997).

The PNS, a cable like nerves extend from CNS to throughout the body. It provides pathways for signals travelling to and from the CNS. This system is made up of millions of nerves cells called neurones. There are three types of neurons, namely the afferent neurons, efferent neurons and inter-neurons. The afferent neurons, also known as sensory or receptor neurons carry impulses from receptors or sensory cells on sense organs to the CNS. On the other hand, the efferent neurons carry impulses from the CNS to the effectors such as muscles. The afferent neurons are always linked to efferent neurons indirectly. In most cases, the efferent neurons reflexion is a reaction to afferent neurons senses. Figure 2.2 shows the connection of both neurons.




Figure 2.2 The connection between afferent neurons and efferent neurons

2.2 Human Sensory Sub-Systems

The human sensory system is a part of human nervous system that responsible for processing sensory information. It consists of sensory receptors that receive stimuli (information) from internal and external environment, neural pathways that conduct the information to brain and parts of brain that involves in sensory perception and process the information. The information is called sensory information and it may lead human to conscious awareness. The interaction of

human with the surrounding environment is carried out through sensory systems that respond to a variety of stimuli such as light, sound, pressure and chemicals.

The human sensory system is an amalgamation of sub-sensing systems, such as

- i) visionary (sight),
- ii) auditory (hearing),
- iii) somatosensory (touch),
- iv) olfaction (smell) and
- v) gustation (taste).

All humans respond to chemical information through a process known as chemoreception. The chemical signals play a major role in feeding, territorial recognition, sexual behaviour and detection of potential harmful conditions such as fire, gas and bad food. Two types of chemoreception have been distinguished which are specialised into smell and taste. These senses normally show the quality of discrimination but sometimes can express the quantitative information.

2.2.1 Olfaction System

The olfaction system represents one of the oldest sensory modalities in the history of mammals (Vokshoor & McGregor, 2010). As a chemical sensor, the olfaction system detects food and influences social and sexual behaviour. It involves the detection and perception of chemical floating in the air called odour.

A simple odour is usually a small, polar molecule that will enter the nasal cavity and then sensed by the human olfaction system. Odour sensations are caused by the interaction of odorant with specialised receptors in the olfaction epithelium in the top of nasal cavity. The signals that are caused by the interaction of odorant with olfaction receptors in the olfaction epithelium are transmitted to the olfaction bulb and finally to the brain.

Odour sensations are usually characterised by the group of description, such as sulphurous, fruity, floral and earthy or by their specific sources such as strawberry or orange. The range of distinguishing odour sensations is wide. A skilled perfume chemist can recognise and distinguish up to 8,000 to 10,000 different substances.

The sense of smell is a sensitive system that can respond to very low concentrations of chemicals. However, it is estimated that only 2% of the volatile compounds available in a single sniff will reach the olfaction receptors and as few as 40 molecules the odorant are sufficient to perceive an odour. In